

EEC Letter of Conformity on Medical Products

We, the manufacturer

Gottlieb Weinmann
Geräte für Medizin und Arbeitsschutz GmbH+Co.
Kronsaaßweg 40 · D - 22525 Hamburg

declare in sole
responsibility the

Product Name: **flowmeter, variable**

Type/Model: **WM 1270, WM 1280
WM 13650, WM 13670**

being in conformity
with the respective
regulations of the
following guideline:

Directive 93/42/EEC on Medical Products

Classification:

II a


Standards used
in particular:

EN 60601-1:1990 + A1:1993 + A2:1995
EN 60601-1-2:1993
DIN EN 738-1

Hamburg, 05. 03. 1998



Dr. J. Gierfahn
Executive President
Chief Executive Officer



Dr. K.-A. Feldhahn
Executive Vice President
Chief Operating Officer

WEINMANN